

Lower Extremity section

**Garment Styles
Design Choices
Modifications
Measuring Guide**

Anklet

0105

Anklet:

For scarring to:

- Below malleolus, not extending to MTP joints
- Can be open or enclosed for toes
- Consider adding a heel lining to prolong life of the garment
- This design is comfortable but consider scar areas carefully so you do not compromise pressure to the required scar area. Make sure garment is 5cm beyond scar area.



Knee Length

0101

For scarring to:

- Lower leg and foot
- Scar should not extend beyond the upper one third of the lower leg
- If scarring extends to the patella or poplitealfossa region – use a full length leg garment as pressure will be compromised
- Open or closed toe styles available



Thigh Length

0201

For scarring to:

- To lower thigh, patella and lower leg
- Silicone elastic can be selected to hold the garment up
- Open or closed toes available
- Zippers may be incorporated into garment if required, however pressure may be slightly compromised
- Consider a chap style one leg as an alternative to prevent garment falling down



Chap Style, with Waistband, One Leg
Chap Style, with Waistband, Two Legs

0035

0036

For scarring to:

- As per thigh length, 0201, but with the addition of a waist band to secure and hold in place.



Foot Glove to Ankle

0538


Foot Glove to Knee

0539

For scarring to:

- Into the MTP JT's, toe webs and individual toes
- Please ensure patient keeps toe nails short, please consider length of these when measuring
- For the 0539 use green tape for leg measurements

A sock with individual toes applying pressure to these area's.



Waist Height, Two Legs, Open Pubis

1102

Waist Height, Two Legs, Closed Pubis

1101

For scarring to:

- To both entire lower extremities,
- If they have been used as donor sites, a garment with bilateral full length legs is suggested.

Note: Scar tissue will mature at a different rate in different areas of the body, this style of garment will probably require revision before the entire leg is mature.



Waist Height, One Legs, Open Pubis

1103

For scarring to:


- Lower abdomen, hips and one leg.

Will not address abdomen scarring therefore consider a panty with short leg

**Chest Height, Two Legs, Pregnancy**

1104

- A reduced pressure panel over the abdomen, which is suitable for pregnant women

**Panty Girdle, Open Pubis**

1122

Panty Girdle, Closed Pubis


1119

For scarring:

- Over the hip joint or the upper thigh
- This type of garment will assure proper pressure over the hip and upper thigh that would be lacking in the panty brief.
- To achieve adequate pressure, the scar should not extend any further than 3" (7.62cm) above the knee.
- Can be constructed with legs of differing lengths and an open or closed pubis.



Panty Girdle (below Knee) Open Pubis	1110
Panty Girdle (below Knee) Closed Pubis	1111



Waist Height, One Leg, Panty, Open Pubis	1112
Waist Height, One Leg, Panty, Closed Pubis	1113

The waist height with one full length leg and the other a short leg

For scarring:

- Over the lower abdomen in addition to one entire leg
- If restricted to the thigh of one leg and the entire length of the opposite leg

Note: consider patients lifestyle, mobility and age when ordering this style.



Panty Brief, Closed Pubis	1120
---------------------------	------


For scarring to:

- To buttocks or pelvic area
- This type of garment should not be used if the burn extends beyond the hip joint or the fold of the buttocks.





Stump

The stump support is measured with the red full length measuring tape, please see diagram.



Waist Height, One Stump	1132
Waist Height, Two Stumps	1133



Waist Height, One Leg plus Stump, Closed Pubis	1134
Waist Height, One Leg plus Stump, Open Pubis	1135

